

## Medical electron accelerators - Part 2: Structural requirements for radiation protection and requirements for testing of approval conditions for operation

---

Contents	Page
Foreword .....	5
1 Scope .....	7
2 Normative references .....	7
3 Terms and definitions and indices .....	7
3.1 Terms and definitions .....	8
3.2 Indices .....	10
4 Dose and dose rate information .....	11
5 Maximum values for the dose behind the shielding .....	11
6 Room requirements .....	11
7 Specifications of the manufacturer and of the radiation protection responsible person ....	12
7.1 Specifications of the manufacturer .....	12
7.2 Specifications of the radiation protection responsible person .....	13
8 Operational data .....	14
8.1 Radiation energy .....	14
8.2 Dose rate .....	14
8.3 Working load WA .....	14
8.4 Direction factor .....	15
8.5 Occupation factor .....	16
8.6 Interaction of factors U and T .....	16
9 Maximum permissible area dose .....	17
10 Rating of radioprotective shielding .....	17
10.1 General .....	17
10.2 General calculation scheme .....	18
10.3 Shielding against bremsstrahlung in the useful radiation direction .....	19
10.4 Shielding against electron useful radiation .....	20
10.5 Shielding against leakage radiation .....	21
10.6 Shielding against secondary bremsstrahlung .....	25
10.7 Shielding at the access to the irradiation room .....	28
10.7.1 Shielding at the radiation protection labyrinth .....	28
10.7.2 Shielding at the gate of an irradiation room without a radiation protection labyrinth .....	29
10.8 Shielding against neutron radiation .....	30
10.9 Shielding the radiation protection labyrinth .....	35
10.10 Interaction of several radiation sources and components .....	37
10.11 Structural radiation protection measures against radioactive material created by nuclear photoprocesses .....	38
11 Constructional radiation protection drawing .....	38
11.1 Specifications of the constructor .....	38
11.2 Safekeeping of the radiation protection drawing .....	39
12 Check for compliance with radiation protection rules .....	39

12.1	General .....	39
12.2	Check of requirements placed on the rooms in accordance with 6 c) to 6 i) .....	39
12.3	Check of the structural shielding .....	39
12.3.1	General .....	39
12.3.2	Requirements on the measuring equipment .....	39
12.3.3	Checking the requirements for the structural shielding .....	40
12.3.4	Evaluation of measured results .....	41
13	Test of ventilation system .....	41
Annex A (informative) Tables for the attenuation curves in Figure 3, Figure 4 and Figure 5 .....		42
Annex B (informative) Calculation example .....		45
Bibliography .....		52
Index of defined terms .....		53
Figures Figure 1 -- Example of the distance $a_n$ for useful radiation .....		19
Figure 2 -- Example of the distance $a_n$ for leakage radiation .....		22
Figure 3 -- Product of the thickness at tenth maximum $z_X$ and density for bremsstrahlung, leakage radiation, X-ray radiation share in the radiation field of the electron radiation, and bremsstrahlung in the useful radiation direction generated by electrons outside the radiation head, for the materials specified in the key .....		23
Figure 4 -- Product of the thickness at tenth maximum $z_X$ and density for bremsstrahlung, leakage radiation, X-ray radiation share in the radiation field of the electron radiation, and bremsstrahlung in the useful radiation direction generated by electrons outside the radiation head, for the materials specified in the key .....		24
Figure 5 -- Factor $k_{XE}$ for calculation of shielding against bremsstrahlung in useful radiation direction generated outside of the radiation head .....		25
Figure 6 -- Example of the distance $a_{Xs}$ for secondary bremsstrahlung .....		27
Figure 7 -- Example of the distance $a_{Xt}$ and surface $A_{Xt}$ for tertiary bremsstrahlung .....		29
Figure 8-- Example of the distance $a_{Nn}$ for direct neutron radiation .....		31
Figure 9 -- Example of distance $a_{Ns}$ for scattered neutron radiation .....		33
Figure 10 -- Labyrinth length and width (schematic) .....		34
Figure B.1 -- Floor plan view for the rating of radioprotective shielding in electron accelerator installations (calculation example) .....		46
Tables Table 1-- Maximum permissible area dose HW .....		17
Table 2 -- Product of thickness at tenth maximum $z_{Xs}$ or $z_{Xt}$ and density for secondary or tertiary bremsstrahlung .....		26
Table 3 -- Thicknesses at tenth maximum $z_{Nn}$ and $z_{Ns}$ for neutron radiation .....		33
Table 4 -- Parameters for rating radioprotective shielding according to 10.2 to 10.9 .....		35
Table 5 -- Reinforcing shielding against the interaction of several radiation components .....		38
Table A.1 -- Product of the thickness at tenth maximum $z_X$ and density for bremsstrahlung, leakage radiation, X-ray radiation share in the radiation field of the electron radiation, and bremsstrahlung in the useful radiation direction generated by electrons outside the radiation head, for the materials listed (see Figure 3) .....		42

Table A.2 -- Product of the thickness at tenth maximum $zX$ and density for bremsstrahlung, leakage radiation, X-ray radiation share in the radiation field of the electron radiation, and bremsstrahlung in the useful radiation direction generated by electrons outside the radiation head, for the materials listed (see Figure 4) .....	43
Table A.3 -- Factor $k_{XE}$ for calculation of shielding against bremsstrahlung in useful radiation direction generated outside of the radiation head (see Figure 5) .....	44
Table B.1 -- Radiation conditions according to specifications of the radiation protection responsible person for the calculation example to rate the radioprotective shielding according to Figure B.1 .....	47
Table B.2 -- Operating data according to specifications of the manufacturer for the calculation example to rate the radioprotective shielding according to Figure B.1 .....	48
Table B.3 -- Calculation data to rate the radioprotective shielding for the various radiation components and occupied areas in the calculation example according to Figure B.1 .....	49