

ISO 8672:2014-05 (E)

Air quality - Determination of the number concentration of airborne inorganic fibres by phase contrast optical microscopy - Membrane filter method

Contents		Page
Foreword		iv
Introduction		v
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	General method description	2
4.1	Limitations of the method by particle type	2
4.2	Limit of visibility and detection limits	2
4.3	Apparatus and equipment	3
4.4	Mounting media	5
4.5	Quality assurance	6
5	Sampling	6
5.1	Flowrate	6
5.2	Acceptable fibre loadings on filters	7
5.3	Storage and transport	7
5.4	Blanks	7
5.5	Sample duration and volume	8
5.6	Sampling strategy and records	9
6	Evaluation	10
6.1	Sample preparation	10
6.2	Microscope adjustment principles	11
6.3	Eyepiece graticule calibration	11
6.4	Microscope/counter performance assessment	11
6.5	Counting and sizing fibres	12
6.6	Calculation of fibre concentration	13
7	Sources of errors and uncertainty	14
7.1	General	14
7.2	Uncertainty	14
7.3	Inter-laboratory variability	16
7.4	Lowering of intra- and inter-laboratory variance	17
8	Test report	17
Annex A (informative) Slide mounting: acetone-triacetin procedure		19
Annex B (informative) Permanent slide preparation: dimethyl formamide - Euparal procedure		21
Annex C (normative) Eyepiece graticule		24
Annex D (informative) Measurement of exposed filter area		27
Annex E (informative) Microscope adjustment procedure		28
Annex F (informative) Sources		29
Bibliography		30