ISO 8672:2014-05 (E)

Air quality - Determination of the number concentration of airborne inorganic fibres by phase contrast optical microscopy - Membrane filter method

Contents

Page

Foreword iv

Introduction v

1 Scope 1

2 Normative references 1

3 Terms and definitions 1

4 General method description 2
4.1 Limitations of the method by particle type 2
4.2 Limit of visibility and detection limits 2
4.3 Apparatus and equipment 3
4.4 Mounting media 5
4.5 Quality assurance 6

5 Sampling 6
5.1 Flowrate 6
5.2 Acceptable fibre loadings on filters 7
5.3 Storage and transport 7
5.4 Blanks 7
5.5 Sample duration and volume 8
5.6 Sampling strategy and records 9

6 Evaluation 10
6.1 Sample preparation 10
6.2 Microscope adjustment principles 11
6.3 Eyepiece graticule calibration 11
6.4 Microscope/counter performance assessment 11
6.5 Counting and sizing fibres 12
6.6 Calculation of fibre concentration 13

7 Sources of errors and uncertainty 14
7.1 General 14
7.2 Uncertainty 14
7.3 Inter-laboratory variability 16
7.4 Lowering of intra- and inter-laboratory variance 17

8 Test report 17

Annex A (informative) Slide mounting: acetone-triacetin procedure 19

Annex B (informative) Permanent slide preparation: dimethyl formamide - Euparal procedure 21

Annex C (normative) Eyepiece graticule 24

Annex D (informative) Measurement of exposed filter area 27

Annex E (informative) Microscope adjustment procedure 28

Annex F (informative) Sources 29

Bibliography 30