ISO 21863:2020 (E)

Water quality — Determination of alkylmercury compounds in water — Method using gas chromatography-mass spectrometry (GC-MS) after phenylation and solvent extraction

Contents

Foreword

Introduction

1 Scope

2 Normative references

3 Terms and definitions

4 Principle

5 Interferences

5.1 Interferences with sampling, sample storage and sample preparation
5.2 Interferences with GC-MS

6 Reagents and standards

7 Apparatus and materials

8 Sample collection, preservation and storage

9 Procedure

9.1 Sample preparation
9.1.1 pH-adjustment of water sample
9.1.2 Phenylation and solvent extraction
9.1.3 Dehydration of toluene extract
9.2 Preparation of samples for GC#MS
9.3 Optimization of operating condition for GC#MS
9.4 Identification of individual substances with GC#MS
9.5 Blank tests

10 Calibration

10.1 General requirements
10.2 Performance test of GC#MS
10.3 Calibration with internal standard
10.3.1 General requirement
10.3.2 Procedure of calibration
10.4 Spike recovery test of target substances

11 Calculation

11.1 Calculation of results after calibration with internal standards
11.2 Treatment of results lying outside the calibration range
11.3 Quality checks for internal standardization

12 Expression of results

13 Test report

Annex A (informative) Example of operating condition of GC-MS

Annex B (informative) Examples of mass chromatograms and mass spectra of phenylated alkylmercury by GC-MS

Annex C (informative) The use of sodium tetrapropylborate as an alternative derivatizing agent[4]

Annex D (informative) The use of GC-AFS as an alternative detector

- **D.1 Principle**
- **D.2 Interferences**
- **D.3 Reagents**
- **D.4 Apparatus and materials**
- **D.5 Procedure**
- **D.6 Calibration**
- **D.7 Calculation**

Annex E (informative) Performance data

Page count: 21