

# ISO 10703:2021 (E)

## Water quality — Gamma-ray emitting radionuclides — Test method using high resolution gamma-ray spectrometry

---

### Contents

	Foreword
	Introduction
1	Scope
2	Normative references
3	Terms and definitions
4	Symbols
5	Principle
6	Reference sources
6.1	Source(s) for energy calibration
6.2	Reference source(s) for efficiency calibration
6.2.1	General
6.2.2	Reference sources for laboratory systems
6.2.3	Reference sources used with numerical methods
7	Reagents
8	Gamma-ray spectrometry equipment
8.1	General description
8.2	Detector types
8.3	High voltage power supply
8.4	Preamplifier
8.5	Cryostat or electric cooler
8.6	Shielding
8.7	Analogue or digital acquisition electronics
8.7.1	General
8.7.2	Analogue electronic (ADC)
8.7.3	Digital electronic (DSP)
8.8	Computer, including peripheral devices and software
9	Nuclear decay data
10	Sampling
11	Procedure
11.1	Sample preparation
11.1.1	General
11.1.2	Direct measurement without preparation
11.1.3	Evaporation without iodine retention
11.1.4	Evaporation with iodine retention
11.2	Calibration
11.2.1	General
11.2.2	Energy calibration
11.2.3	Efficiency calibration
12	Expression of results
12.1	Calculation of the activity concentration

- 12.1.1 General
  - 12.1.2 Dead time and pile up corrections (see ISO 20042)
  - 12.1.3 Decay corrections
  - 12.1.4 True coincidence summing
    - 12.1.4.1 General
    - 12.1.4.2 Probability of true-coincidence summing
    - 12.1.4.3 Consequences for source-based efficiency calibrations and for sample count
    - 12.1.4.4 Reducing the effect of true-coincidence summing
  - 12.2 Standard uncertainty
  - 12.3 Decision threshold
  - 12.4 Detection limit
  - 12.5 Limits of the coverage intervals
    - 12.5.1 Limits of the probabilistically symmetric coverage interval
    - 12.5.2 The shortest coverage interval
  - 12.6 Corrections for contributions from other radionuclides and background
    - 12.6.1 General
    - 12.6.2 Contribution from other radionuclides
    - 12.6.3 Contribution from background
- 13 Test report
- Annex A (informative) Example of a carrier solution which can be added to the water sample when waste water from a nuclear power plant is investigated
- Annex B (informative) True coincidence summing
  - B.1 Gamma-gamma coincidence
  - B.2 Gamma-x-ray coincidence
  - B.3 Gamma-annihilation photon coincidence
- Annex C (informative) Calculation of the activity concentration from a gamma spectrum using a linear background subtraction (undisturbed peak)

Page count: 27