

DIN 19266:2015-05 (E)

pH measurement - Reference buffer solutions for the calibration of pH measuring equipment

Contents		Page
Foreword		3
1	Scope	4
2	Normative references	4
3	Terms and definitions	4
4	Fundamentals	5
4.1	General	5
4.2	pH values of primary reference buffer solutions	5
4.3	pH values of secondary reference buffer solutions	6
5	Preparation of primary and secondary reference buffer solutions	9
5.1	General	9
5.2	Reference buffer solution A, pH(S) = 1,679 (25 °C), Potassium tetraoxalate	9
5.3	Reference buffer solution B, pH(S) = 3,557 (25 °C), Potassium hydrogen tartrate	9
5.4	Reference buffer solution C, pH(S) = 4,005 (25 °C), Potassium hydrogen phthalate	9
5.5	Reference buffer solution D, pH(S) = 6,865 (25 °C), Phosphate	9
5.6	Reference buffer solution E, pH(S) = 7,413 (25 °C), Phosphate	9
5.7	Reference buffer solution F, pH(S) = 9,180 (25 °C), Borax	9
5.8	Reference buffer solution G, pH(S) = 12,454 (25 °C), Calcium hydroxide	10
5.9	Reference buffer solution H, pH(S) = 3,776 (25 °C), Potassium dihydrogen citrate	10
5.10	Reference buffer solution I, pH(S) = 10,012 (25 °C), Sodium carbonate/sodium hydrogen carbonate	10
6	Storage and shelf life	10
7	Examples of pH(S) values of reference buffer solutions	10
8	Additional properties of reference buffer solutions	10
Annex A (informative)	Examples of pH(S) values as a function of temperature at temperatures from 55 °C to 95 °C	12
Annex B (informative)	Volumetric method	13
Bibliography		15