ISO 17280:2015-05 (E)

Surface active agents - Determination of 1,4-dioxan residues in surfactants obtained from epoxyethane by gas chromatography

Contents

Page

Foreword iv

1 Scope 1

2 Normative references 1

3 Principle 1

4 Reagents 1
4.1 1,4-Dioxane, purity 99,5 % minimum 1
4.2 4-Methyl-1,3-dioxane, purity 98 % minimum 1
4.3 N, N-dimethylformamide (DMF) 1
4.4 Carrier gas: nitrogen or helium, purity 99,99 % minimum 2

5 Apparatus 2
5.1 Analytical balance, 0,1 mg readability 2
5.2 Gas chromatography system, attached with temperature programmed device and flame ionization detector (FID) 2
5.3 Head space sampling unit 2
5.4 20 ml crimped top glass vials, sealing pliers and sealing clamp suitable for head space unit 2
5.5 Septum and crimped tops, lined with aluminium or PTFE, for closing the vials 2
5.6 Sealing pliers, for the vials 2
5.7 Capillary column, capable of the separation characteristics shown in Figure 2 2
5.8 Volumetric flasks, 50 ml and 100 ml capacity 2
5.9 Gas tight syringes or glass pipettes, 1 ml ± 0,01 ml capacity 2

6 Preparation of calibration solutions 2
6.1 Calibration solution A 2
6.2 Calibration solution B 3

7 Procedure 3
7.1 Preparation of the test sample 3
7.2 Standard addition procedure 3
7.2.1 Preparation of test sample solutions 3
7.2.2 Determination 4
7.2.3 Expression of results 4
7.3 Relative response factor procedure 4
7.3.1 Determination of relative response factor 4
7.3.2 Preparation of test sample solutions 5
7.3.3 Determination 5
7.3.4 Expression of results 5

8 Chromatographic conditions 5

9 Evaluation 6
9.1 General 6
9.2 Identification 7
9.3 Calculation 7

10 Precision 7

10.1 Repeatability limit 7
10.2 Reproducibility limit 7

11 Test report 7

Annex A (informative) Statistical data from the ring carried out 8