

ISO 18337:2015-06 (E)

Surface chemical analysis - Surface characterization - Measurement of the lateral resolution of a confocal fluorescence microscope

Contents		Page
Foreword		iv
Introduction		v
1	Scope	1
2	Terms and definitions	1
3	Symbols and abbreviated terms	1
4	General	1
4.1	Background information	1
4.2	Types of CFM operation	2
4.2.1	General	2
4.2.2	Stage scanning CFM	2
4.2.3	Laser scanning CFM	2
4.2.4	Spinning disk CFM	2
4.3	Parameters that affect the lateral resolution of a CFM	3
4.3.1	General	3
4.3.2	Objective lens	3
4.3.3	Detection pinhole size and focal length of the tube lens	3
4.3.4	Collimation and purity of the laser illumination beam	3
4.3.5	Polarization of the laser illumination	3
4.3.6	Excitation and emission wavelengths	3
4.3.7	Image contrast	3
5	Measuring the lateral resolution by imaging a small object	3
5.1	Background information	3
5.2	Selection of the sample and sample requirements	4
5.3	Setting the parameters prior to operating the instrument	4
5.4	Data collection and analysis	5
5.4.1	Selecting a proper spot	5
5.4.2	Extracting a line profile using band-average process	5
5.5	Recording the data	6
Annex A (informative)	Sample preparation and example of data and analysis	7
Bibliography		10