

# DIN EN ISO 22232-1:2021-09 (E)

## Non-destructive testing - Characterization and verification of ultrasonic test equipment - Part 1: Instruments (ISO 22232-1:2020)

---

<b>Contents</b>		<b>Page</b>
<b>European foreword</b> .....		<b>4</b>
<b>Foreword</b> .....		<b>5</b>
<b>1</b>	<b>Scope</b> .....	<b>6</b>
<b>2</b>	<b>Normative references</b> .....	<b>6</b>
<b>3</b>	<b>Terms and definitions</b> .....	<b>6</b>
<b>4</b>	<b>Symbols</b> .....	<b>8</b>
<b>5</b>	<b>General requirements of conformity</b> .....	<b>9</b>
<b>6</b>	<b>Manufacturer's technical specification for ultrasonic instruments</b> .....	<b>9</b>
<b>7</b>	<b>Performance requirements for ultrasonic instruments</b> .....	<b>12</b>
<b>8</b>	<b>Group 1 tests</b> .....	<b>14</b>
8.1	Equipment required for group 1 tests .....	14
8.2	Battery operational time .....	14
	8.2.1 Procedure .....	14
	8.2.2 Acceptance criterion .....	15
8.3	Stability after warm-up time .....	15
	8.3.1 Procedure .....	15
	8.3.2 Acceptance criteria .....	15
8.4	Stability against temperature .....	15
	8.4.1 Procedure .....	15
	8.4.2 Acceptance criterion .....	17
8.5	Stability against voltage variation .....	17
	8.5.1 Procedure .....	17
	8.5.2 Acceptance criterion .....	18
8.6	Time base deviation .....	18
	8.6.1 Procedure .....	18
	8.6.2 Acceptance criterion .....	20
8.7	Transmitter pulse parameters .....	20
	8.7.1 General .....	20
	8.7.2 Pulse repetition frequency .....	20
	8.7.3 Effective output impedance .....	20
8.8	Receiver .....	21
	8.8.1 General .....	21
	8.8.2 Cross talk from transmitter to receiver during transmission .....	21
	8.8.3 Dead time after transmitter pulse .....	22
	8.8.4 Dynamic range and maximum input voltage .....	24
	8.8.5 Receiver input impedance .....	25
	8.8.6 Time-corrected gain (TCG) .....	26
8.9	Gates .....	27
	8.9.1 General .....	27
	8.9.2 Gates with value output .....	28
	8.9.3 Gates with analogue output .....	30
	8.9.4 Gates with alarm output .....	32
8.10	Highest digitized frequency .....	33
	8.10.1 Procedure .....	33

	8.10.2	Acceptance criterion.....	34
8.11		Response time of digital ultrasonic instruments.....	34
	8.11.1	General.....	34
	8.11.2	Procedure.....	34
	8.11.3	Acceptance criterion.....	35
<b>9</b>		<b>Group 2 tests.....</b>	<b>35</b>
9.1		Equipment required for group 2 tests.....	35
9.2		Physical state and external aspects.....	36
	9.2.1	Procedure.....	36
	9.2.2	Acceptance criterion.....	36
9.3		Transmitter voltage, pulse rise time and duration.....	36
	9.3.1	Procedure.....	36
	9.3.2	Acceptance criteria.....	39
9.4		Receiver.....	39
	9.4.1	General.....	39
	9.4.2	Frequency response.....	39
	9.4.3	Noise level.....	41
	9.4.4	Gain linearity.....	42
	9.4.5	Vertical display linearity.....	42
		<b>Annex A (normative) Special conditions for ultrasonic instruments with logarithmic amplifiers.....</b>	<b>44</b>
		<b>Bibliography.....</b>	<b>45</b>