

DIN 53000-2:2023-07 (E)

Viscometry - Measurement of kinematic viscosity by means of the Ubbelohde viscometer - Part 2: Sources of errors and corrections

Contents

Page

Foreword	3
1 Scope	4
2 Normative references	4
3 Terms and definitions.....	4
4 Symbols and units	5
5 Principle	6
6 Basic principles.....	7
7 Errors of measurement that can be corrected.....	7
7.1 Kinetic energy correction (“Hagenbach correction”).....	7
7.2 Correction for surface tension	8
7.3 Correction for air buoyancy.....	9
7.4 Correction for difference in pressure heads	9
7.5 Temperature correction	10
7.6 Correction of the thermal expansion of the capillaries and the timing bulb	10
7.7 Correction for volume expansion	10
7.8 Correction for inclination.....	11
7.9 Correction of the spatial dependence of the acceleration due to gravity	11
8 Errors of measurement that cannot be corrected	11
8.1 Time measurement.....	11
8.2 Temperature control	12
8.3 Heat due to friction.....	12
8.4 Turbulence	12
8.5 Trailing film	13
9 Errors of measurement caused by random influences	13

Tables

Table 1 — Symbols and units	5
Table 2 — Relative errors of measurement ε_V as a percentage when the connecting line between the centres of the suspended level bulbs deviates from the perpendicular ($\varphi_0 = 0$) by the angle φ	11