

ISO 15632:2021-02 (E)

Microbeam analysis - Selected instrumental performance parameters for the specification and checking of energy-dispersive X-ray spectrometers (EDS) for use with a scanning electron microscope (SEM) or an electron probe microanalyser (EPMA)

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	3
4.1 General description	3
4.2 Energy resolution	4
4.3 Dead time	4
4.4 Peak-to-background ratio	4
4.5 Energy dependence of instrumental detection efficiency	5
5 Check of further performance parameters	5
5.1 General	5
5.2 Stability of the energy scale and resolution	5
5.3 Pile-up effects	5
5.4 Periodical check of spectrometer performance	5
Annex A (normative) Measurement of line widths (FWHMs) to determine the energy resolution of the spectrometer	6
Annex B (normative) Measurement of the L/K ratio as a measure for the energy dependence of the instrumental detection efficiency	11
Bibliography	13