

# ISO 4962:2024-11 (E)

## Nanotechnologies - In vitro acute nanoparticle phototoxicity assay

---

<b>Contents</b>		<b>Page</b>
Foreword .....		v
Introduction .....		vi
1	Scope .....	1
2	Normative references .....	1
3	Terms, definitions, symbols and abbreviated terms .....	1
3.1	Terms and definitions .....	1
3.2	Abbreviated terms .....	2
4	Test method .....	2
5	Materials and equipment .....	3
5.1	Materials .....	3
5.1.1	Reagents .....	3
5.1.2	Cell line .....	3
5.1.3	Controls .....	4
5.2	Apparatus .....	5
5.2.1	Laminar flow cabinet, standard biological hazard .....	5
5.2.2	Incubator (37 °C, 95 % humidified, 5 % CO <sub>2</sub> /air) .....	5
5.2.3	Inverted phase contrast microscope .....	5
5.2.4	Centrifuge .....	5
5.2.5	Water bath .....	5
5.2.6	Multiple well plate reader .....	5
5.2.7	Tissue culture flasks .....	5
5.2.8	24 multi-well plates with flat bottom .....	5
5.2.9	Flat bottom 96-well black polystyrene microplate .....	5
5.2.10	8-channel pipette, 20-200 l Hemocytometer .....	5
5.2.11	Conical tube .....	5
5.2.12	Reservoir .....	5
5.2.13	Vortex mixer .....	5
5.2.14	Refrigerator .....	5
5.2.15	Freezer .....	5
5.2.16	UV crosslinker (capable of light emission at =365 nm), as a UVA source. Irradiation of the test plate should be between 4 and 6 mW/cm <sup>2</sup> .....	5
5.2.17	UV-VIS spectrophotometer, capable of measurements in the wavelength range from 300 to 800 nm .....	5
5.2.18	UV power meter, capable of measurements in the wavelength range from 315 to 400 nm ...	5
6	Sample preparation .....	5
7	Preparations .....	6
7.1	General .....	6
7.2	Culture medium .....	6
7.3	Preparation of cell stock culture .....	6
7.4	Verify viable cell growth .....	6
7.5	Irradiation conditions .....	8
7.5.1	UVA source .....	8
7.5.2	Light dose (insolation) measurement .....	8
7.6	Multiple well plate reader .....	8

<b>8</b>	<b>Measurement procedure .....</b>	<b>8</b>
<b>8.1</b>	<b>Cell seeding (Day 1) .....</b>	<b>10</b>
<b>8.2</b>	<b>Incubation of cells with the positive control and NP suspension (Day 2) .....</b>	<b>10</b>
<b>8.3</b>	<b>UVA exposure (Day 3) .....</b>	<b>11</b>
<b>8.4</b>	<b>Cell viability assay (Day 4) .....</b>	<b>11</b>
<b>8.5</b>	<b>Evaluation of artefacts due to possible NP interferences with the MTS assay (in the dark) .....</b>	<b>12</b>
<b>8.6</b>	<b>Data analysis .....</b>	<b>13</b>
<b>9</b>	<b>Report .....</b>	<b>14</b>
<b>9.1</b>	<b>Test report .....</b>	<b>14</b>
<b>9.2</b>	<b>Report data format .....</b>	<b>15</b>
<b>10</b>	<b>Precision .....</b>	<b>15</b>
<b>10.1</b>	<b>Repeatability .....</b>	<b>15</b>
<b>10.2</b>	<b>Reproducibility .....</b>	<b>15</b>
	<b>Annex A (informative) Schematic diagram of 96-well plate position .....</b>	<b>16</b>
	<b>Annex B (informative) Verification of plate reader uniformity .....</b>	<b>17</b>
	<b>Annex C (informative) Dispersing procedure for TiO<sub>2</sub> nanoparticles in DMEM .....</b>	<b>18</b>
	<b>Annex D (informative) Results of the inter-laboratory test .....</b>	<b>19</b>
	<b>Bibliography .....</b>	<b>22</b>