

ISO 16666:2025-11 (E)

Surface chemical analysis - Total reflection X-ray fluorescence - Principles and general requirements

Contents		Page
Foreword		iv
Introduction		v
1	Scope	1
2	Normative references	1
3	Terms, definitions and abbreviations	1
3.1	Abbreviated terms	1
4	Physical background of TXRF	2
4.1	General	2
4.2	Definition of angle and X-ray path in TXRF geometry	3
4.3	Beam projection	4
4.4	X-ray standing waves (XSW)	4
4.5	Fluorescence intensity	6
5	Instrumental requirements	6
5.1	Description of various possible TXRF measurement setups	6
5.2	Beam conditioning	8
5.2.1	Requirement	8
5.2.2	Beam guidance	8
5.2.3	Beam modulation (tuning)	8
5.2.4	Measurement angles in TXRF	8
5.3	Reflector (sample carrier)	9
5.3.1	General requirements	9
5.3.2	Reflectors for chemical analysis	9
5.4	Reflector alignment	9
5.5	Detector	10
5.6	Sample station	10
5.7	Control unit	10
6	Calibration and quality control	10
6.1	Angle inspection	10
6.2	Energy calibration	10
6.3	Calibration of the element sensitivities	11
6.4	Inspection of the sample excitation	12
6.5	Inspection of the element detection	12
7	Spectra analysis and Reporting	12
7.1	Spectra Processing	12
7.2	Reporting Spectra	13
8	Samples	13
9	Qualitative and quantitative analysis	14
9.1	Qualitative analysis	14
9.2	Quantitative analysis	14
9.2.1	Methods of quantitative analysis	14
9.3	Preparation of a measurement for quantitative analysis	15

9.4	Procedure of a quantitative analysis by the element-specific sensitivity method (internal standard addition)	16
9.5	Verification of quantitative TXRF measurements	17
9.6	Traceability in TXRF	17
Annex A (informative)	Measurement angles in TXRF, background correction, peak correction, spectrum deconvolution	18
Annex B (informative)	Round-robin test for verification of the precision and accuracy of TXRF spectroscopy	21
Bibliography	22