

# ISO 4266-3:2002-11 (E)

## Petroleum and liquid petroleum products - Measurement of level and temperature in storage tanks by automatic methods - Part 3: Measurement of level in pressurized storage tanks (non-refrigerated)

---

<b>Contents</b>		<b>Page</b>
<b>1</b>	<b>Scope .....</b>	<b>1</b>
<b>2</b>	<b>Normative reference .....</b>	<b>1</b>
<b>3</b>	<b>Terms and definitions .....</b>	<b>1</b>
<b>4</b>	<b>Precautions .....</b>	<b>3</b>
<b>4.1</b>	<b>Safety precautions .....</b>	<b>3</b>
<b>4.2</b>	<b>Equipment precautions .....</b>	<b>3</b>
<b>4.3</b>	<b>General precautions .....</b>	<b>3</b>
<b>5</b>	<b>Accuracy .....</b>	<b>4</b>
<b>5.1</b>	<b>Intrinsic error of ALGs .....</b>	<b>4</b>
<b>5.2</b>	<b>Calibration prior to installation .....</b>	<b>4</b>
<b>5.3</b>	<b>Error caused by installation and operating conditions .....</b>	<b>4</b>
<b>5.4</b>	<b>Overall accuracy .....</b>	<b>4</b>
<b>5.4.1</b>	<b>General .....</b>	<b>4</b>
<b>5.4.2</b>	<b>Use of ALGs for fiscal/custody transfer applications .....</b>	<b>5</b>
<b>6</b>	<b>Installation of ALGs .....</b>	<b>5</b>
<b>6.1</b>	<b>General .....</b>	<b>5</b>
<b>6.2</b>	<b>Mounting location .....</b>	<b>5</b>
<b>6.3</b>	<b>Manufacturer's requirements .....</b>	<b>5</b>
<b>6.4</b>	<b>Installation .....</b>	<b>5</b>
<b>6.4.1</b>	<b>Installation of intrusive ullage ALGs (e.g. float-operated and servo-operated type) mounted on still-wells .....</b>	<b>5</b>
<b>6.4.2</b>	<b>Installation of intrusive ullage ALGs (e.g. float and servo-operated types) using guide wires .....</b>	<b>5</b>
<b>6.4.3</b>	<b>Installation of non-intrusive ullage ALGs (e.g. microwave or radar) mounted on still-wells .....</b>	<b>9</b>
<b>6.4.4</b>	<b>Location of ALG .....</b>	<b>9</b>
<b>6.5</b>	<b>Still-well design .....</b>	<b>9</b>
<b>7</b>	<b>Initial setting and initial verification of ALGs in the field .....</b>	<b>10</b>
<b>7.1</b>	<b>Preparation .....</b>	<b>10</b>
<b>7.1.1</b>	<b>Check for critical distances .....</b>	<b>10</b>
<b>7.1.2</b>	<b>Check for free movement of level-sensing element of intrusive ALGs .....</b>	<b>10</b>
<b>7.1.3</b>	<b>Check for influences of changes in physical and electrical properties of the product (vapour/liquid) .....</b>	<b>10</b>
<b>7.2</b>	<b>Initial setting .....</b>	<b>10</b>
<b>7.2.1</b>	<b>General .....</b>	<b>10</b>
<b>7.2.2</b>	<b>Setting of intrusive ullage ALGs .....</b>	<b>11</b>
<b>7.2.3</b>	<b>Setting of non-intrusive ALGs .....</b>	<b>11</b>
<b>7.3</b>	<b>Initial field verification .....</b>	<b>12</b>
<b>7.3.1</b>	<b>Introduction .....</b>	<b>12</b>
<b>7.3.2</b>	<b>Verification procedure .....</b>	<b>12</b>
<b>7.3.3</b>	<b>Tolerance of initial field verification .....</b>	<b>13</b>
<b>7.4</b>	<b>Record keeping .....</b>	<b>13</b>

<b>8</b>	<b>Subsequent verification of ALGs .....</b>	<b>13</b>
<b>8.1</b>	<b>General .....</b>	<b>13</b>
<b>8.2</b>	<b>Frequency of subsequent verification .....</b>	<b>13</b>
<b>8.3</b>	<b>Procedure .....</b>	<b>13</b>
<b>8.4</b>	<b>Tolerance for subsequent verification of ALGs in fiscal/custody transfer application .....</b>	<b>13</b>
<b>8.5</b>	<b>Comparison of the ALG readings from current and previous verifications .....</b>	<b>14</b>
<b>8.6</b>	<b>Adjustment in subsequent verification .....</b>	<b>14</b>
<b>9</b>	<b>Data communication and receiving .....</b>	<b>14</b>
	<b>Bibliography .....</b>	<b>15</b>