

DIN EN ISO 20417:2021-12 (E)

Medical devices - Information to be supplied by the manufacturer (ISO 20417:2021)

Contents

Page

European foreword	4
Foreword	5
Introduction	6
1 Scope	7
2 Normative references	7
3 Terms and definitions	8
4 General considerations	15
5 Information elements to be established	16
5.1 Units of measurement	16
5.2 Graphical information	16
5.3 Language and country identifiers	17
5.3.1 Language identifiers	17
5.3.2 Country identifiers	17
5.4 Dates	17
5.5 Full address	18
5.6 Commercial product name	18
5.7 Model number	18
5.8 Catalogue number	18
5.9 Production controls	18
5.10 Unique device identifier	19
5.11 Types of use/reuse	19
5.12 Sterile	19
6 Requirements for accompanying information	19
6.1 Requirements for information to be supplied on the label	19
6.1.1 Minimum requirements for the label	19
6.1.2 Identification of the manufacturer	20
6.1.3 Identification of the medical device or accessory	21
6.1.4 Other label requirements	23
6.1.5 Consult instructions for use	24
6.1.6 Safety signs	25
6.2 Identification requirements for detachable components of a medical device or accessory	26
6.3 Legibility of the label	26
6.4 Durability of markings	26
6.5 Information to be provided on the packaging	27
6.5.1 General information	27
6.5.2 Packaging for the lay user	28
6.5.3 Special conditions indicated on the packaging	29
6.6 Requirements for information in the instructions for use and technical description	30
6.6.1 General	30
6.6.2 Requirements for instructions for use	31
6.6.3 Additional requirements for the instructions for use for a lay user	36
6.6.4 Requirements for technical description	36
6.6.5 Requirements for e-documentation	39

7	Other information that is required to be supplied with the <i>medical device</i> or <i>accessory</i>	39
7.1	<i>Importer</i>	39
7.2	<i>Distributor</i>	39
7.3	Repackaging	40
7.4	Translation	40
7.5	Regulatory identification	41
	Annex A (informative) Particular guidance and rationale	42
	Annex B (informative) Example test method for assessing <i>clearly legible</i> requirements	45
	Annex C (informative) Example test method for assessing durability	46
	Annex D (informative) Cross reference between the document and the requirements considered	47
	Annex E (informative) Reference to the IMDRF <i>essential principles</i> and labelling guidances	59
	Annex F (informative) Reference to the <i>essential principles</i>	63
	Annex G (informative) Reference to the general safety and performance requirements for <i>medical devices</i>	67
	Annex H (informative) Reference to the general safety and performance requirements for <i>IVD medical devices</i>	71
	Annex I (informative) Terminology — Alphabetized index of defined terms	75
	Bibliography	77